



Quality assurance of quantitative cardiac T1-mapping in multicenter clinical trials – A T1 phantom program from the hypertrophic cardiomyopathy registry (HCMR) study



Qiang Zhang^{a,*}, Konrad Werys^a, Iulia A. Popescu^a, Luca Biasioli^a, Ntobeko A.B. Ntusi^b, Milind Desai^c, Stefan L. Zimmerman^d, Dipan J. Shah^e, Kyle Autry^e, Bette Kim^f, Han W. Kim^g, Elizabeth R. Jenista^g, Steffen Huber^h, James A. Whiteⁱ, Gerry P. McCann^j, Saidi A. Mohiddin^k, Redha Boubertakh^l, Amedeo Chiribiri^m, David Newbyⁿ, Sanjay Prasad^o, Aleksandra Radjenovic^p, Dana Dawson^q, Jeanette Schulz-Menger^r, Heiko Mahrholdt^s, Iacopo Carbone^t, Ornella Rimoldi^u, Stefano Colagrande^v, Linda Calistri^v, Michelle Michels^w, Mark B.M. Hofman^x, Lisa Anderson^y, Craig Broberg^z, Flett Andrew^{aa}, Javier Sanz^{ab}, Chiara Bucciarelli-Ducci^{ac}, Kelvin Chow^{ad}, David Higgins^{ae}, David A. Broadbent^{af}, Scott Semple^{ag}, Tarik Hafyane^{ah}, Joanne Wormleighton^{ai}, Michael Salerno^{aj}, Taigang He^{ak}, Sven Plein^{al}, Raymond Y. Kwong^{am}, Michael Jerosch-Herold^{an}, Christopher M. Kramer^{ao}, Stefan Neubauer^a, Vanessa M. Ferreira^a, Stefan K. Piechnik^a

^a Oxford Centre for Clinical Magnetic Resonance Research, Oxford BRC NIHR, Division of Cardiovascular Medicine, Radcliffe Department of Medicine, University of Oxford, UK

^b Department of Medicine, University of Cape Town and Groote Schuur Hospital, Cape Town, South Africa

^c Cleveland Clinic, USA

^d Johns Hopkins Hospital, USA

^e Houston Methodist DeBakey Heart & Vascular Center, USA

^f Mount Sinai West Hospital; Icahn School of Medicine at Mount Sinai, USA

^g Duke Cardiovascular Magnetic Resonance Center, Duke University Medical Center, USA

^h Department of Radiology and Biomedical Imaging, Yale School of Medicine, New Haven, USA

ⁱ Stephenson Cardiac Imaging Centre, Libin Cardiovascular Institute, University of Calgary, Canada

^j Department of cardiovascular sciences, University of Leicester and NIHR Leicester Biomedical Research Centre, UK

^k Inherited Cardiovascular Diseases, Barts Heart Centre, London, UK

^l Barts Heart Centre, St Bartholomew's Hospital, Barts Health NHS Trust, West Smithfield, London, UK

^m King's College London and Guy's and St Thomas' NHS Foundation Trust, UK

ⁿ Centre for Cardiovascular Science, University of Edinburgh, UK

^o National Heart and Lung Institute, Imperial College and Royal Brompton Hospital, London, UK

^p Institute of Cardiovascular & Medical Sciences, BHF Glasgow Cardiovascular Research Centre, University of Glasgow, Glasgow, UK

^q Aberdeen Cardiovascular and Diabetes Centre, College of Life Sciences and Medicine, University of Aberdeen, UK

^r Charité, University Medicine Berlin ECRC and Helios Clinics, Berlin, Germany

^s Department of Cardiology, Robert Bosch Medical Center, Stuttgart, Germany

^t Department of Radiological, Oncological and Pathological Sciences, Sapienza, University of Rome, Italy

^u Università Vita Salute San Raffaele, Milan, Italy

^v Department of Experimental and Clinical Biomedical Sciences, University of Florence, Italy

^w Erasmus MC, department of cardiology, Rotterdam, the Netherlands

^x dept. Radiology and Nuclear Medicine, Amsterdam UMC location VUmc, Amsterdam, The Netherlands

^y Cardiology Clinical Academic Group, St George's University of London, UK

^z Knight Cardiovascular Institute, Oregon Health and Science University, USA

^{aa} University Hospital Southampton NHS Trust, UK

^{ab} Mount Sinai Hospital, New York, NY, USA

^{ac} University of Bristol, UK

^{ad} Siemens Medical Solutions USA, Inc., Chicago, IL, USA

^{ae} Philips Electronics UK Limited, Surrey, UK

^{af} Biomedical Imaging Sciences Department, University of Leeds, Leeds, UK

^{ag} Edinburgh Imaging, Centre for Cardiovascular Science, University of Edinburgh, UK

^{ah} Montreal Heart Institute, Canada

^{ai} University Hospitals of Leicester NHS Trust, UK

^{aj} University of Virginia, USA

^{ak} The Cardiology Clinical Academic Group (CAG), St George's University of London, St George's University Hospitals NHS Foundation Trust, UK

* Corresponding author.

E-mail address: qiang.zhang@cardiov.ox.ac.uk (Q. Zhang).

^{a1} Department of Biomedical Imaging Science, Leeds Institute of Cardiovascular and Metabolic Medicine, University of Leeds, UK^{am} Cardiovascular Division, Department of Medicine, Brigham and Women's Hospital, Harvard Medical School, USA^{an} Brigham and Women's Hospital, USA^{ao} University of Virginia Health, USA

ARTICLE INFO

Article history:

Received 28 November 2020

Accepted 7 January 2021

Available online 31 January 2021

Keywords:

Cardiac MRI

Quantitative T1-mapping

Multicenter study

Quality assurance

Standardization

Phantom study

ABSTRACT

Background: Quantitative cardiovascular magnetic resonance T1-mapping is increasingly used for myocardial tissue characterization. However, the lack of standardization limits direct comparability between centers and wider roll-out for clinical use or trials.

Purpose: To develop a quality assurance (QA) program assuring standardized T1 measurements for clinical use.

Methods: MR phantoms manufactured in 2013 were distributed, including ShMOLLI T1-mapping and reference T1 and T2 protocols. We first studied the T1 and T2 dependency on temperature and phantom aging using phantom datasets from a single site over 4 years. Based on this, we developed a multiparametric QA model, which was then applied to 78 scans from 28 other multi-national sites.

Results: T1 temperature sensitivity followed a second-order polynomial to baseline T1 values ($R^2 > 0.996$). Some phantoms showed aging effects, where T1 drifted up to 49% over 40 months. The correlation model based on reference T1 and T2, developed on 1004 dedicated phantom scans, predicted ShMOLLI-T1 with high consistency (coefficient of variation 1.54%), and was robust to temperature variations and phantom aging. Using the 95% confidence interval of the correlation model residuals as the tolerance range, we analyzed 390 ShMOLLI T1-maps and confirmed accurate sequence deployment in 90%(70/78) of QA scans across 28 multiple centers, and categorized the rest with specific remedial actions.

Conclusions: The proposed phantom QA for T1-mapping can assure correct method implementation and protocol adherence, and is robust to temperature variation and phantom aging. This QA program circumvents the need of frequent phantom replacements, and can be readily deployed in multicenter trials.

© 2021 The Author(s). Published by Elsevier B.V. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

1. Introduction

CMR parametric mapping permits the quantification and spatial visualization of changes in myocardial composition based on changes in T1, T2 relaxation times, and extracellular volume (ECV) [1]. Native T1-mapping has been shown to have narrow normal ranges within the same method, and sensitivity to a wide range of myocardial diseases [2,3]. Its advantages include relatively simple single breath-hold acquisitions, excellent reproducibility, and avoidance of gadolinium-based contrast agents.

However, T1 values depend on the protocol parameters, acquisition and reconstruction methods [4], currently requiring cumbersome establishment of within-center norms [1,5]. Native T1-mapping can differentiate disease from healthy tissue in single-site studies at a fixed field strength, and demonstrated reproducibility in multi-center studies under a uniform imaging set-up, but the lack of standardization limits direct comparability between centers and wider roll-out for clinical use. For clinical sites attempting to implement T1 mapping, it is often unclear how to install and validate the methods correctly before using them for clinical diagnosis, or for assessing novel therapeutics in multi-center trials. Despite SCMR consensus recommendations [1], there has been no working solution to-date for standardization – not even for a single T1-mapping technique on a single vendor platform. The use of local normal ranges and z-scores to adjust for differences have been proposed [1], but are suboptimal, as small sample size (e.g. 10–20) healthy volunteers can be prone to sampling errors [1,5]; outliers or biases could directly affect a center's ability to detect abnormal findings to diagnose disease. While these effects decrease with larger sample sizes, re-acquiring such data frequently just to monitor stability after each method or scanner upgrade becomes prohibitively cumbersome and expensive.

Validation and quality assurance (QA) of single-method deployment could be a first step to standardize T1-mapping techniques, increasing the confidence of individual centers in the set-up of CMR mapping for clinical use. The large, international multicenter Hypertrophic Cardiomyopathy Registry (HCMR) study [6] adopted a single T1-mapping method (Shortened Modified Look-Locker Inversion

Recovery, ShMOLLI [7]) to maximize comparability of the datasets to power for outcomes. This study setup provided an opportunity to collect the required phantom data to develop a QA program for standardizing T1-mapping between centers, using an original multiparametric modelling approach. The derived QA model can reliably detect deviations from correct method implementation and protocol adherence, despite changes in phantom properties, including temperature variations [8,9] and aging effects, unlike most phantom solutions currently available [9,10].

2. Materials and methods

2.1. Study design

In 2013, a batch of 50 original dedicated QA phantom devices was designed for the prototype ShMOLLI T1-mapping method [7]. The manufacture details are provided in Supplemental Section S1. The external appearance and arrangement of the phantom compartments are shown in Fig. 1a, b, and the achieved T1 and T2 combinations in Table 1.

The QA program aimed to assure the detection and compensation of potential differences in T1-mapping sequence properties between sites to track T1 measurement stability [6]. Centers operating Siemens MR scanners and performing measurements free of charge ($N = 28$) were provided a consistent protocol developed at Oxford core lab, adapted for various Siemens software platforms. The sites were provided with imaging manuals and HCMR QA protocols to perform repeated ShMOLLI-T1 acquisitions, an inversion recovery spin echo (IR-SE) acquisition and a multi-echo SE acquisition for reference T1 and T2 maps (protocol specified in Supplemental Section S2). The multi-center dataset collected from the HCMR sites served to monitor site-specific changes of the T1-mapping technique properties. Meanwhile phantoms were repeatedly scanned at Oxford core lab on a MAGNETOM Avanto (1.5 T) and a MAGNETOM Trio Tim (3 T) scanners (Siemens Healthcare, Erlangen, Germany). This high-volume dataset served to investigate temperature and age dependencies as well as to establish the QA protocol.

The phantom scans were uploaded by individual HCMR sites to Boston core lab and then sent to Oxford core lab for data analysis between July 2014 and December 2017, when the first phase of the HCMR study was completed [6]. 28 HCMR sites using Siemens scanners acquired and sent 78 scans (a complete list of HCMR sites providing the phantom scans is given in Supplemental Table S1). Meanwhile, locally at Oxford core lab, we acquired 441 phantom studies between October 2013 and June 2017. Each study scanned one to three phantoms, which provided a total of 1004 phantom scans for analysis.

All reference T1 and T2 maps were calculated offline by a C++ library [11]. We analyzed the T1 and T2 using the robust mean values from the circular (15 mm




Fig. 1. HCMR QA phantom. (a) Phantom external appearance. (b) Phantom compartment arrangement. (c) An example of T1 map in QA post-processing. Black dashed lines indicate the automatically detected ROIs. The two phantoms on either side are not part of QA; their use is recommended to assure adequate coil loading.

diameter) regions of interest within each of the phantom compartments (Fig. 1c), which were placed automatically by a machine learning algorithm [12]. Altogether, from each set of phantom scans, we obtained 9 sets of reference T1 ($T1_{ref}$) and T2 measurements paired with 5 repeats of corresponding ShMOLLI-T1 ($T1_{sh}$) for each compartment.

2.2. T1 and T2 dependencies on temperature and phantom aging

To investigate the T1 and T2 dependency on temperature, we scanned phantoms at temperatures ranging from 16 °C to 28 °C. To analyze phantom property drifts over time, four selected phantoms were scanned approximately every two weeks between February 2014 and June 2017, with an average of 148 studies acquired per phantom over a 40-month period. We evaluated time trends in $T1_{ref}$ and T2, corrected to room temperature (21 °C) using the established temperature dependencies.

2.3. T1 prediction model and T1-mapping quality assurance

We postulated that the measured cardiac ShMOLLI-T1 has a predictable dependency on $T1_{ref}$ and T2 in phantoms. Using the dataset acquired locally at Oxford core lab, we established a single empirical equation exploiting the inherent relationship, $T1_{sh} = \text{Function}(T1_{ref}, T2)$, for all T1 and T2 combinations. The residual of fitting was calculated to establish the range of agreement for the QA program. The equation was then used to predict the expected ShMOLLI-T1 from $T1_{ref}$ and T2, and compare it with observed ShMOLLI-T1 in the same QA scan to check the agreement. Phantom scans with ShMOLLI-T1 of all compartments within the agreement range (the 95% confidence interval (CI) of the residual of fitting; see results section) would indicate an accurate ShMOLLI sequence deployment and, therefore, QA passed. If any ShMOLLI-T1 fell outside the agreement range, the pattern of discrepancy and the scanning parameters were then further inspected, to identify potential sources of error. These may include artefacts, reconstruction error or protocol error, and are given a conditional pass with specific recommendations for remedial actions. Failing to identify and rectify errors would suggest that the ShMOLLI sequence was deployed incorrectly, and, therefore, QA failed. The capacity of this method was validated on the multi-center HCMR phantom dataset for Siemens scanners, with proof-of-principle translation demonstrated on the Philips platform.

Table 1

Phantom chemical composition and T1/T2 relaxation times measured at room temperature within one month of manufacture. ShMOLLI T1 are average values of a batch of 50 phantoms measured at 1.5 T and 3 T. The T2 are average values of five randomly selected phantoms at 1.5 T and another five at 3 T scanned in an additional single measurement.

Phantom compartment and formulation	ShMOLLI T1 [ms] (mean \pm SD, N = 50)		Spin-echo T2 [ms] (mean \pm SD, N = 5)	
	1.5 T (21.3 \pm 0.4 °C)	3 T (21.0 \pm 0.5 °C)	1.5 T (21.0 °C)	3 T (21.0 °C)
A 0.5% Agar, 0.33% Carrageenan, 0.113 mM NiCl ₂	2529.5 \pm 14.0	2461.0 \pm 7.8	275.8 \pm 2.1	265.9 \pm 4.8
B 0.5% Agar, 0.626 mM NiCl ₂	1396.2 \pm 5.9	1329.4 \pm 2.2	266.2 \pm 4.6	259.0 \pm 2.2
C Undoped 18 M Ω deionized H ₂ O	3251.5 \pm 12.5	3234.9 \pm 27.8	2373.4 \pm 184.3	2383.1 \pm 153.9
D 1.9% Agar, 1.2 mM NiCl ₂	859.1 \pm 2.9	804.71 \pm 0.74	72.2 \pm 0.7	71.4 \pm 0.4
E 2% Agar, 0.77 mM NiCl ₂	1109.7 \pm 12.5	1051.3 \pm 1.9	69.6 \pm 1.2	70.7 \pm 0.8
F 2% Agar, 0.524 mM NiCl ₂	1397.8 \pm 4.7	1328.0 \pm 2.2	80.2 \pm 1.4	78.4 \pm 2.8
G 1.5% Agar, 0.1% Carrageenan, 4.5 mM NiCl ₂	323.1 \pm 0.66	307.5 \pm 0.6	72.4 \pm 0.7	68.5 \pm 1.1
H 3% Agar, 0.457 mM NiCl ₂	1428.8 \pm 3.9	1368.1 \pm 3.4	56.9 \pm 0.4	56.5 \pm 2.6
I 1.8% Agar, 2 mM NiCl ₂	610.2 \pm 1.3	574.4 \pm 0.5	74.0 \pm 0.7	72.5 \pm 1.5

3. Results

3.1. T1 baseline values and temperature dependency

At the time of manufacture, all 9 compartments of the 50 phantoms made as a batch had consistent ShMOLLI-T1 values with coefficient of variation within 0.85%, all measured at room temperature of \sim 21 °C (Table 1). In contrast to in-vivo myocardial T1 values [7], ShMOLLI-T1 values of all phantom compartments were lower at 3 T than at 1.5 T by 0.5–6.8% (all $p < 0.001$). Phantom T2 were typically lower at 3 T by 2.1–3.7% (all $p < 0.001$), with the exception of compartment #E where T2 was 1.6% higher at 3 T than at 1.5 T ($p < 0.001$). In compartments #C and #H, the T2 difference did not reach statistical significance (Table 1).

In subsequent experiments, $T1_{ref}$ showed clear linear dependency on temperature at both 1.5 T and 3 T (Fig. 2a). The variation of predicted T1 changes with temperature was $2.6\% \pm 1.5\%$ at 1.5 T and $2.6\% \pm 1.2\%$ at 3 T for all compartments, relative to the baseline T1 in Table 1. The temperature sensitivity increased with baseline T1, following closely a second-order polynomial ($R^2 > 0.996$; Fig. 2b).

3.2. Impact of phantom aging on T1

A wide range of aging effects were observed on the reference T1 measurements, with the largest seen in compartment #B in three of the four phantoms scanned repeatedly at Oxford core lab (Fig. 3). These phantoms appear to have undergone a transition, whereby the baseline T1 values had increased by \sim 50%. This may be due to various factors during the manufacture, particularly the amount of air within




Fig. 2. Temperature sensitivity of reference T1 in the Oxford core lab dataset at 1.5 T and 3 T. (a) T1 temperature dependency. Temperature sensitivity coefficients ($\Delta T1/\Delta t$, ms/°C) are provided to the right of the graph, prefixed with the compartment ID. Regression lines are omitted for clarity. (b) T1 temperature sensitivity coefficients (Y-axis) follow a second-order polynomial to baseline T1 values (X-axis).

the universal container tubes and the resilience of the caps. T2 characteristics are provided in supplemental Fig. S4 and S5.

3.3. T1 prediction model robust to temperature variation and phantom aging

Consistent with prior work, ShMOLLI-T1 generally followed a linear relationship with T1ref, with visibly larger underestimation in compartments with shorter T2 (Fig. 4a, red arrows). To establish a model independent of phantom age and temperature, we exploited the inherent relationships between ShMOLLI-T1 and T1ref and T2, based on the 1004 dedicated phantom datasets. We accounted for the linear dependency of ShMOLLI-T1 on T1ref, and an exponential dependency on T2 (coefficients calculated for 1.5 T and 3 T, respectively; full data samples in Supplemental Fig. S6). While this model significantly reduced the residual differences between ShMOLLI-T1 and reference T1 methods from 7.24% to 1.64% ($p < 0.001$), there was a remaining small trend visible in the fit residuals. This was rectified with a third-order polynomial correction. The final multivariate correlation model predicting the ShMOLLI-T1 ($\widehat{T1}_{sh}$) from reference measurements were thus established; see derived equations in Supplemental S3. The model predicted $\widehat{T1}_{sh}$ which agreed with real T1sh with high accuracy ($R^2 > 0.99$, Fig. 4b), despite the temperature variation and phantom aging effects described above. Analysis of the residual errors of the ShMOLLI prediction model allowed the establishment of the 95% confidence interval $CI = \pm 3.12\%$, and the 99.7% confidence interval $CI = \pm 5.32\%$, robust to temperature and aging confounders. The 95% CI was used as the tolerance range for the QA (full data samples in Supplemental Fig. S7).

3.4. Clinical application: T1-mapping quality assurance

We received 94 phantom scans: 78 from 28 Siemens sites, 15 from 6 Philips sites, and 1 from a General Electric (GE) site (complete list of sites in Supplemental Table S1). The vendor and scan distribution reflected the resources available and local feasibility during this study, which precluded fair head-to-head inter-vendor comparisons. We were able to perform QA and present the findings in 78 scans (390 ShMOLLI-T1 maps) from the 28 Siemens sites (Table 2 and Fig. 5). We exemplified the need for further work for inter-vendor application of the QA model (Fig. 5f), subject to sufficient datasets acquired on those MR systems.

3.5. QA passed

QA was passed if the dataset from a session contained at least one accurate ShMOLLI-T1 acquisition for all 9 compartments. 34 scans from 15 sites gave ShMOLLI-T1 values within the prescribed 95% CI range of the expected ShMOLLI-T1, confirming accurate ShMOLLI deployment (Fig. 5a). Three scans from 3 sites showed departure of ShMOLLI-T1 values in individual phantom compartments, with the rest being within the agreement range. We identified the sources of these outliers as imaging artefacts in the reference T1-maps (Fig. 5b). These 3 scans were considered to have passed the QA, as the source of discrepancies was clearly outside the cardiac T1-mapping technology, and the ShMOLLI-T1 values of the rest of the compartments were within the agreement range.




Fig. 3. The varied appearance of age-related drifts in reference T1 (T_{1ref}) in individual phantom compartments observed over a period of 40 months. All T_{1ref} values are corrected to room temperature. Colors represent the four phantoms investigated.

3.6. Common error 1: Lower T_1 s due to inadequate waiting time

A further 26 scans from 16 sites provided at least one accurate ShMOLLI acquisition but with a characteristic pattern of greater underestimation of ShMOLLI- T_1 values in longer T_1 compartments in other acquisitions of the scan (Fig. 5c), suggesting incomplete recovery of magnetization. This was likely caused by inadequate waiting time after the previous image acquisition, frequency adjustment, or shimming adjustment. We reproduced this T_1 underestimation at Oxford core lab with dedicated experiments, and studied the impact of incomplete magnetization recovery on the measured T_1 . The experimental details are provided in Supplemental Section S4. This error may be rectified by reminding the operators to wait at least 10 s in between T_1 -map acquisitions.

3.7. Common error 2: failed in-line reconstruction of T_1 -mapping

We detected incorrect in-line ShMOLLI- T_1 reconstruction in 7 additional scans from 5 sites (Fig. 5d, gray circles). We were able to reconstruct and salvage T_1 -maps offline using raw T_1 -weighted images, and restored the accurate T_1 values (Fig. 5d, blue points). This indicated correct T_1 -weighted image acquisition but inaccurate inline reconstruction on the MR system. Therefore, these sites passed QA conditionally on offline reconstruction, but required redeployment of the T_1 sequence.

3.8. QA failed

Six scans did not provide T_1 measurements falling wholly within the prescribed tolerance limit, and therefore did not pass QA (Fig. 5e). A further 2 scans from 2 sites contained no reference T_1 or T_2 acquisitions, and thus the QA was not performed. Similarly, incompatible

measurement protocols prevented a convincing inter-vendor validation, with one example shown as a proof of principle (Fig. 5f).

4. Discussion

In this work, we have established a novel approach for quality assurance (QA) that is independent of the actual physical properties of the phantoms, bypassing the exhibited phantom sensitivities to temperature variations and aging. We have demonstrated how to use this QA to verify correct implementation of T_1 -mapping methods to within a prescribed tolerance range across multiple scanners and magnetic field settings; signature patterns of departure also identified common errors for actionable remediations.

4.1. Inter-method and inter-vendor applications

We have illustrated how to achieve this phantom QA solution using a single T_1 -mapping method, but the solution can be deployed to other mapping techniques. The HCMR Consortium had chosen a single T_1 -mapping method (ShMOLLI [7]) to maximize comparability of the datasets to power for outcomes. The ShMOLLI T_1 -mapping sequence possesses three characteristics which made this work possible within 5 years: (1) method stability and full accountability of its quantitative characteristics, such that it was feasible to compare it on a wide range of platforms within a single vendor setting; (2) heart-rate independence, alleviating the need to vary heart rate as part of the QA process; (3) ShMOLLI is a single universal technique suitable for measuring a wide range of T_1 values (whether short or long), thus obviating the need for developing separate QA models, each for a specific MOLLI sub-variant for short or long T_1 s.










Fig. 4. ShMOLLI T1 (T1sh) against reference T1 (T1ref) in the Oxford core lab dataset before T2 correction (a) and after T2 correction (b) at 1.5 T and 3 T. (a) The visible deviations (red arrows) were driven predominantly by T2 effects. (b) The multivariate correlation model including T2 effects predicts T1sh with high accuracy ($R^2 > 0.99$).

Table 2

QA results of 78 Siemens scans from 28 sites with proposed outcomes and actions.

QA Results	Description	No. scans (sites)	Action recommended
1. Passed	All T1 maps in the scan provided ShMOLLI-T1 (T1sh) values within the agreement range with expected $\hat{T}1sh$ (Fig. 5a). Disagreement between T1sh and $\hat{T}1sh$ in one or more individual compartments; the rest were in the agreement range. The source of error can be identified as image artefacts not linked to cardiac T1 maps (Fig. 5b).	34 scans (15 sites) 3 scans (3 sites)	QA passed. No further action required QA passed. Consider technical investigation.
2. Warnings	Underestimated T1sh in individual acquisitions. At least one acquisition is within the agreement range. Incomplete recovery of longitudinal magnetization in individual acquisitions (Fig. 5c).	26 scans (16 sites)	QA passed, with warning of possible protocol adherence problems
3. Conditional	T1sh values outside the agreement range; Source of disagreement caused by T1 map fitting without ShMOLLI conditional fitting reconstruction, but accurate T1sh values were successfully restored offline (Fig. 5d).	7 scans (5 sites)	QA conditional on offline reconstruction. Require re-deployment of T1 sequence
4. Failed	T1sh values outside the agreement range; unable to identify source of error. Unable to restore accurate T1sh values offline (Fig. 5e). QA could not be performed due to missing reference T1 or T2 sequences in the scan.	6 scans (3 sites) 2 scans (2 sites)	QA not passed. Technical investigation required Incomplete scan. Check protocols and repeat QA



The QA solution may be applied to any other quantitative mapping methods across vendor platforms, subject to the exact application of the specified protocols. The provided scanning protocol (Supplemental Section S2) and presented QA formula can be readily tested on other vendor MR systems as published, subject to the availability of vendor research personnel and support of technical development sites. Even though our experience indicates that this is far from trivial, by setting clear and achievable QA goals, the proposed QA program offers for the first time a practical way towards a fully compatible inter-vendor T1 methodology.

4.2. Study limitations and future work

The proposed QA program was demonstrated on a single vendor platform. Although the general approach can be applied across vendors, the investigators' experience indicates that translation to other vendors and methods will require significant further technical investment, to address the large intervendor differences between documented similar methods [10]. The authors are open to collaborative requests to employ the proposed QA to improve the T1-mapping consistency in multi-vendor setting. In future, the described QA should be either implemented directly on the scanner or be provided as a prompt external service, so that end-users can assure method stability with the same frequency as the usual scanner QA. This QA is designed to test correct technical deployment of methods; operator compliance with the acquisition protocols and standardization of image processing [13] are required to minimize introduction of errors.

5. Conclusions

We presented the development of a practical MR phantom QA program for accurate and comparable T1 measurements for use in a large multi-center setting. The QA model is robust to phantom aging and ambient temperature variations, circumventing the need for manual temperature corrections and frequent phantom replacements. This provides an immediate and economical solution to verify correct T1-mapping method implementation, and identify common errors for remediation. The proposed QA program paves the way to widespread adoption of T1-mapping into routine clinical practice.

Funding

This work is funded by British Heart Foundation (BHF) project grant PG/15/71/31731 (QZ, KW, SKP) and National Institute for Health Research (NIHR) Oxford Biomedical Research Centre at The Oxford University Hospitals (IAP, SKP, VMF and SN).

Declaration of Competing Interest

SKP has patent authorship rights for U.S. patent US20120078084A1. Systems and methods for shortened Look Locker inversion recovery (Sh-MOLLI) cardiac gated mapping of T1. Granted March 15, 2016. IP is managed by Oxford University Innovations; the license exclusively transferred to Siemens Healthcare.

QZ, SKP, KW, IAP, VMF have authorship rights for pending patent PCT/GB2020/051189. A method for identity validation and quality assurance of quantitative magnetic resonance imaging protocols. Filed

May 15, 2020. IP is owned and managed by Oxford University Innovations.

Acknowledgements

This research was made possible by the consistent support and encouragement by the HCMR steering committee, to proceed with this study without additional reimbursement for additional scanning time. We would like to acknowledge Yuna Choi and Kathleen Cheng at the Boston HCMR core lab, and Matthias Friedrich at McGill University for their support; Aaron Long for data transfer within OCMR, and Karen Whatley at OCMR for packing and sending the phantoms.

Appendix A. Supplementary data

Supplementary data to this article can be found online at <https://doi.org/10.1016/j.ijcard.2021.01.026>.

References

- [1] D.R. Messroghli, J.C. Moon, V.M. Ferreira, L. Grosse-Wortmann, T. He, P. Kellman, et al., Clinical recommendations for cardiovascular magnetic resonance mapping of T1, T2, T2* and extracellular volume: a consensus statement by the Society for Cardiovascular Magnetic Resonance (SCMR) endorsed by the European Association for Cardiovascular Imaging (EACVI), *J. Cardiovasc. Magn. Reson.* 19 (1) (2017) 75.
- [2] E.B. Schelbert, D.R. Messroghli, State of the art: clinical applications of cardiac T1 mapping, *Radiology* 278 (3) (2016) 658–676.
- [3] J.M. Liu, A. Liu, J. Leal, F. McMillan, J. Francis, A. Greiser, et al., Measurement of myocardial native T1 in cardiovascular diseases and norm in 1291 subjects, *J. Cardiovasc. Magn. Reson.* 19 (2017) 74.
- [4] I.A. Popescu, K. Werys, Q. Zhang, H. Puchta, E. Hann, E. Lukaschuk, et al., Standardization of T1-mapping in cardiovascular magnetic resonance using clustered structuring for benchmarking normal ranges, *Int. J. Cardiol.* (2020).
- [5] J.C. Moon, D.R. Messroghli, P. Kellman, S.K. Piechnik, M.D. Robson, M. Ugander, et al., Myocardial T1 mapping and extracellular volume quantification: a Society for Cardiovascular Magnetic Resonance (SCMR) and CMR Working Group of the European Society of Cardiology consensus statement, *J. Cardiovasc. Magn. Reson.* 15 (1) (2013) 92.
- [6] C.M. Kramer, E. Appelbaum, M.Y. Desai, P. Desvigne-Nickens, J.P. DiMarco, M.G. Friedrich, et al., Hypertrophic cardiomyopathy registry: the rationale and design of an international, observational study of hypertrophic cardiomyopathy, *Am. Heart J.* 170 (2) (2015) 223–230.
- [7] S.K. Piechnik, V.M. Ferreira, E. Dall'Armellina, L.E. Cochlin, A. Greiser, S. Neubauer, et al., Shortened Modified Look-Locker Inversion recovery (ShMOLLI) for clinical myocardial T1-mapping at 1.5 and 3 T within a 9 heartbeat breathhold, *J. Cardiovasc. Magn. Reson.* 12 (2010) 69.
- [8] W.D. Zech, N. Schwendener, A. Persson, M.J. Warntjes, C. Jackowski, Temperature dependence of postmortem MR quantification for soft tissue discrimination, *Eur. Radiol.* 25 (8) (2015) 2381–2389.
- [9] G. Captur, P. Gatehouse, K.E. Keenan, F.G. Heslinga, R. Bruehl, M. Prothmann, et al., A medical device-grade T1 and ECV phantom for global T1 mapping quality assurance—the T-1 mapping and ECV standardization in cardiovascular magnetic resonance (TIMES) program, *J. Cardiovasc. Magn. Reson.* 18 (2016) 58.
- [10] G. Captur, A. Bhandari, R. Brühl, B. Ittermann, K.E. Keenan, Y. Yang, et al., T(1) mapping performance and measurement repeatability: results from the multi-national T (1) mapping standardization phantom program (TIMES), *J. Cardiovasc. Magn. Reson.: Off. J. Soc. Cardiovasc. Magn. Reson.* 22 (1) (2020) 31.
- [11] K. Werys, I. Dragonu, Q. Zhang, I. Popescu, E. Hann, H. Puchta, et al., Total mapping toolbox (TOMATO): an open source library for cardiac magnetic resonance parametric mapping, *SoftwareX* 11 (2020) 100369.
- [12] L. Biasioli, J.A. Noble, M.D. Robson, Multicontrast MRI registration of carotid arteries in atherosclerotic and normal subjects, *Medical Imaging 2010: Image Processing*, International Society for Optics and Photonics, 2010.
- [13] V. Carapella, H. Puchta, E. Lukaschuk, C. Marini, K. Werys, S. Neubauer, et al., Standardized image post-processing of cardiovascular magnetic resonance T1-mapping reduces variability and improves accuracy and consistency in myocardial tissue characterization, 298, 2020 128–134.

Fig. 5. Phantom QA of T1-mapping for 6 example CMR centers. In each panel, the top graph shows observed ShMOLLI T1 (T1sh) and the bottom graph the residuals (y-axis) displayed against the expected $\widehat{T1sh}$ (x-axis). (a) QA passed as all within 95% CI (green range). (b) QA passed with artefacts in individual compartment(s). (c) QA passed with warning. Underestimated T1sh in individual acquisitions. (d) QA conditional on T1sh offline reconstruction. Inline reconstruction failed (gray circles). (e–f) QA failed due to presence of patterns and variability of the observed residuals reaching outside the tolerance range.